
Clavibacter michiganense pv. Michiganense (Cmm) Rapid Test

Cat.No: DTSPB1

Lot. No. (See product label)

Size

50 Tests

Intended Use

The test can be simply used to on-site and rapid detect Cmm , the causal agent of bacterial canker of tomato, in infected host plants, without requiring any special equipment or expertise to run.

General Description

Cmm is an economically important pathogen that is seed transmitted and mainly infects tomato, causing tomato canker, infected tomato seeds give rise to contaminated seedlings. Seed testing is an essential tool for the control of Cmm and is important for its regulation and control through phytosanitary certification and quarantine programs in the domestic and international seed trade.

Principle of The Test

The development of rapid test strips is based on Lateral Flow Device (LFD) technology using specific antibodies. A LFD format consists of antibody-coated latex beads latex (or colloidal gold) which will bind the specific pathogen antigen absorbed from the plant extract. The agglutination is accumulated at a specific location by the presence of a line of target specific antibody, which immobilises the agglutinated latex whilst allowing the background reagents to be washed away by continued flow along the membrane. The membrane contains a strip of target-specific antibody, test strip and a strip of another antibody that binds to the specific antibody, control strip. Latex beads containing bound antigen are trapped on the test strip leaving a visible line. Surplus latex beads that do not contain antigen are trapped on the control strip showing that the assay has worked. This provides a non-subjective and clear read out of a positive result against a low background.

Reagents and Materials Provided

1. 50 single-use test cassettes
2. 50 pre-filled extraction tubes with sample buffer
3. 50 disposable pipettes for sample application

Storage

- 1) Store the strips and the extraction buffer at 4°C. However, exposure to ambient temperature (10 - 30°C), such as during transport and use in the field, does not affect the quality. Keep the packaging (containing desiccant bags) always hermetically closed. Absorbed moisture by the strips can lead to poor results or even complete failure of the test.
- 2) Strips must be used before the expiration date indicated on the label of the packaging.

Assay Procedure

- 1) Remove a sample from infected root /leaf, add to the extraction bag.
- 2) Add contents of bottle to the extraction bag. Use the extraction tool to macerate the sample.
- 3) Draw some of the liquid from the extraction bag.
- 4) Add 2 drops to the sample well on the lateral flow device.
- 5) Wait until the colored control line appears and read the test (5 minutes).

Interpretation of Results

Observed the result within 5min, if it presents two red zones and determined as positive, if it only presents one red control plot zone and determined as negative. If the reaction line and the control line all present evenly unknown, or only the reaction line presents red zone and the control line presents nothing ,then determined the reagent strip invalid.

Performance Characteristics

Please Note:
Test kits and cassettes are not reusable.

Sensitivity

The Rapid Test Strip provides high sensitivity comparable with ELISA

Precautions

Please Note:
Test kits and cassettes are not reusable.